

## Sample Statistical Summary

MAPEP-12-XaW27

Analyte	T(1)	A(2)	Grand(3) Mean	Std Dev	Ref Value	Ref Unc	Acceptance Range
Radiological							Units: (Bq/L)
Iodine-129	26	25	6.40	0.64	6.82	0.10	4.77 - 8.87

**Note:** (1) T = Total number of laboratories reporting analyte.  
 (2) A = Number of laboratories with 'Acceptable' performance.  
 (3) Mean excludes values derived as total metals and values indicated as "Not Acceptable".

### Other Flags:

A = Result acceptable Bias  $\leq 20\%$   
 W = Result acceptable with warning  $20\% < \text{Bias} < 30\%$   
 N = Result not acceptable Bias  $> 30\%$   
 L = Uncertainty potentially too low (for information purposes only)  
 H = Uncertainty potentially too high (for information purposes only)  
 RW = Report Warning  
 NR = Not Reported